

Combination
Calibration Source
Wavelength Meter




Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
532 mn	2 Mhz	WS8-2	ND:YAG	± 100 kHz
	10 MHz	WS8-10	ND:YAG	± 100 kHz
	30 Mhz	WS7-30	ND:YAG	± 100 kHz
	60 Mhz	WS7-60	ND:YAG	± 100 kHz

Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
633 mn	2 Mhz	WS8-2	I2-HeNe	± 11 kHz
	10 MHz	WS8-10	I2-HeNe	± 11 kHz
	30 Mhz	WS7-30	I2-HeNe	± 11 kHz
			SL03	± 5 MHz
	60 Mhz	WS7-60	I2-HeNe	± 11 kHz
			SL03	± 5 MHz
			SL02	± 25 MHz

Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
780 mn	2 Mhz	WS8-2	SLR 780	± 2 MHz
	10 MHz	WS8-10	SLR 780	± 2 MHz
	30 Mhz	WS7-30	SLR 780	± 2 MHz
	60 Mhz	WS7-60	SLR 780	± 2 MHz

Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
852 mn	2 Mhz	WS8-2	SLR 852	± 2 MHz
	10 MHz	WS8-10	SLR 852	± 2 MHz
	30 Mhz	WS7-30	SLR 852	± 2 MHz
	60 Mhz	WS7-60	SLR 852	± 2 MHz



Combination
Calibration Source
Wavelength Meter



Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
895 nm	2 Mhz	WS8-2	SLR 895	± 2 MHz
	10 MHz	WS8-10	SLR 895	± 2 MHz
	30 Mhz	WS7-30	SLR 895	± 2 MHz
	60 Mhz	WS7-60	SLR 895	± 2 MHz

Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
1064 nm	2 Mhz	WS8-2	ND:YAG	± 100 kHz
	10 MHz	WS8-10	ND:YAG	± 100 kHz
	30 Mhz	WS7-30	ND:YAG	± 100 kHz
	60 Mhz	WS7-60	ND:YAG	± 100 kHz

Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
1532 nm	2 Mh	WS8-2	SLR 1532	± 2 MHz
	10 MHz	WS8-10	SLR 1532	± 2 MHz
	30 Mhz	WS7-30	SLR 1532	± 2 MHz
	60 Mhz	WS7-60	SLR 1532	± 2 MHz

Wavelength	Accuracy	Wavelength Meter	Calibration Source	Frequenzy stability
3392 nm	3000 Mhz	WS5 IR-III	IR-HeNe	± 75 MHz
	20000 MHz	LSA IR-III	IR-HeNe	± 75 MHz



HighFinesse GmbH
Wöhrdstraße 4
72072 Tübingen, Germany



T + 49 (0) 7071 - 53918 0
F + 49 (0) 7071 - 5391899
M info@highfinesse.com



Additional information
and distributors:
www.highfinesse.com